NeuroSigma

MONARCH
eTNS System

External Trigeminal Nerve Stimulation system for
Attention Deficit Hyperactivity Disorder **Rx Only**



Caution: Federal law restricts this device
to sale by or on the order of a physician

CONTENTS

Labeling Definitions and Symbol Descriptions ..... page 4

Contraindications, Warnings and Precautions ..... page 6

Indications For Use ..... page 9

Trigeminal Nerve Stimulation Therapy ..... page 9

General Description ..... page 10

Monarch® Key Functions ..... page 12

Operating Procedures ..... page 14

Before Treatment ..... page 14

Treatment ..... page 18

Course of Treatment ..... page 19

After Treatment ..... page 21

Please read this manual carefully before beginning treatment. It is an essential part of the Monarch® eTNS® System. This manual should be saved for future reference.

Maintenance and Cleaning ..... page 22

Environmental Specifications ..... page 24

Manufacturer's Declaration – Electromagnetic Emissions ..... page 24

Manufacturer's Declaration – Electromagnetic Immunity ..... page 25

Recommended Separation Distance ..... page 27

Electrical Specifications ..... page 28

Electrical Stimulation ..... page 29

Environmental Conditions ..... page 29

Clinical Data ..... page 31

Warranty ..... page 34

Caution: Federal law restricts this device to sale by or on the order of a physician.

Labeling Definitions and Symbol Descriptions



Caution: Pay special attention to the following details or consult accompanying documents



Type BF Applied Part



By order of a physician (not applicable in the EU)



Manufacturer



Temperature limitations



Use by date

Labeling Definitions and Symbol Descriptions (continued)



Serial number



Do not reuse



Refer to instruction manual



Operating instructions



Product number



Lot number



Unsafe within the magnetic resonance environment



Recycle product separate from other disposables



Keep Dry



Humidity Limitation



Atmospheric Pressure Limitation

IP22 Degree of Protection

Contraindications, Warnings and Precautions

READ ALL INSTRUCTIONS BEFORE USING

⚠ Contraindications

The device is contraindicated for use by patients with:

- implanted cardiac and/or neurostimulation systems
- implanted metallic or electronic device in their head

Do not use the Monarch eTNS System:

- in children under the age of 7 years
- in patients with body worn devices (e.g. insulin pumps and t-VNS)
- within 100cm of RFID readers or emitters that operate below 150kHz
- in the presence of the following sources of radio frequency (RF) energy:
  - MRI
  - Diathermy
  - Cell phones
  - CT
  - Microwave or shortwave
  - Metal detectors
  - RFID tags
  - Electronic article surveillance systems
  - TV transmission lines
  - High frequency (HF) equipment (e.g. surgical cauterizing tools and welding equipment)

Do not attach the electric patches:

- anywhere on the body other than the forehead
- on the neck
- on the chest
- over a defect in the skull (i.e. post brain surgery or head trauma) or over an open fontanel

⚠ Adverse Events

eTNS is generally well-tolerated. However, during the eTNS pivotal study, the following events occurred at a higher rate in the group treated with the eTNS system compared to subjects treated with the sham (placebo) device. Patients should consult their physician if any of the following events are observed:

- Bronchitis
- Headache
- Itching
- Lightheadedness
- Nausea
- Poor appetite
- Skin rash
- Stomach ache
- Tooth pain
- Vomiting
- Trouble sleeping
- Nightmares
- Drowsiness
- Fatigue
- Tingling
- Rapid heartbeat
- Constipation
- Frequent urination
- Increased appetite
- Clenching teeth

Contraindications, Warnings and Precautions (continued)

⚠ Warnings

Children 7 to 12 years receiving eTNS treatment should be closely supervised by an adult who has read the user manual and is familiar with the Monarch eTNS System.

The Monarch eTNS System should:

- only be used by the individual for whom it is intended.
- be used with the guidance of a licensed physician.
- be used with caution in patients with heart disease or serious medical disorders.
- be kept out of the reach of infants and children under the age of 7 years.
- be used only as directed and be applied to healthy, clean, intact skin.
- not be used with other electronic therapeutic devices.
- not be applied on the neck or chest.
- not be used in the presence of electric monitoring equipment (e.g. cardiac monitors).
- not be used in the bath or shower.
- not be used while operating machinery.

The Monarch electric patches should not:

- be used in patients with dermatitis or sensitive skin, as they are at higher risk of developing irritation.
- be used over broken, infected, or inflamed skin.
- be removed carelessly as this may damage the skin.

The Monarch lead wires should not be allowed to wrap around the neck.

The safety of the Monarch is unknown in the following conditions:

- trigeminal neuralgia or injury to the trigeminal nerve
- pregnancy
- concurrent use of hearing aids

The long-term effects of using the Monarch eTNS System are unknown. The effectiveness of the Monarch for long-term use in the indicated population (i.e. for more than 4 weeks) has not been systematically evaluated in controlled trials. Therefore, the physician who elects to use the Monarch for extended periods in patients with ADHD should periodically re-evaluate the long-term usefulness of the device for the individual patient.

Indications For Use

The Generation 2.0 Monarch external Trigeminal Nerve Stimulation (eTNS) System® ("Monarch Gen 2") is indicated for treatment of pediatric Attention Deficit Hyperactivity Disorder (ADHD) as a monotherapy in patients ages 7 through 12 years old who are not currently taking prescription ADHD medications. The device is used for patient treatment by prescription only and is intended to be used in the home under the supervision of a caregiver during periods of sleep.

Trigeminal Nerve Stimulation Therapy

The main role of the trigeminal nerve is to convey sensation information from the face into the brain. External trigeminal nerve stimulation (eTNS) uses this natural pathway to send therapeutic signals into the brain. The trigeminal nerve connects directly from the face to the brainstem, and from there information is sent to other key brain regions thought to be involved in medical conditions such as ADHD (alternatively called hyperkinetic disorder or HKD). While the exact mechanism of action of eTNS is not yet known, neuroimaging studies have shown that eTNS increases activity in brain regions that are known to be important in regulating attention, emotion, and behavior. These findings are supplemented by basic science studies in animals, as well as clinical studies in humans. For more information on eTNS, please visit www.monarch-etns.com.


General Description


The Monarch eTNS System is designed to deliver electrical stimulation to branches of the trigeminal nerve, located just below the skin above the eyebrows, via a cutaneous electric patch.

Main component

Monarch (external pulse generator)

1. OLED Screen
2. + button (increase intensity)
3. 0 button (pause/resume)
4. - button (decrease intensity)
5. USB port (charging)
6. on/off button
7. lock/unlock slider
8. Stereo Port (therapy/patch port)





Standard accessories and spare parts

Electric patches (4 pouches: 28 patches)

Lead wires (1)

USB-C charging cable (1)

AC adapter wall charging plug (1)

User Manual (1)

⚠ Use only NeuroSigma approved accessories and spare parts. Use of unapproved accessories may cause personal injury, loss of therapeutic effect, equipment damage and will void manufacturer's warranty.

Monarch Key Functions

Control buttons

Pressing the + button will increase the intensity level of the treatment. Each momentary press will increase the intensity by 0.1 mA.

Pressing the - button will decrease the intensity level of the treatment. Each momentary press will decrease the intensity by 0.1 mA.

The central O button is used to pause and resume the treatment. A flashing display indicates that the treatment is paused. The exterior timer ring sweeps in 30second intervals, with blue indicating a treatment cycle, and grey indicating a rest cycle. The interior ring counts down the 8-hour treatment duration



Lock slider

Moving this slider to the locked position will disable your controls, and prevent accidental button press in the night. The padlock icon on your display screen will appear closed.

To unlock, move the slider to the unlocked position . On screen, your padlock icon will appear open.

ON/OFF button

This button turns the power on and off. Press and hold for 5 seconds.

Before Treatment

Plug Monarch device into the wall charger, using the included USB-A to USB-C cable. Plug the wall charger into a household socket (AC mains).

The display will indicate that the device is charging if the device is powered on by pressing and holding the power button for 5 seconds. If the device is not powered on, the device will still charge but the screen will not indicate it.

Charging the device completely can take up to four hours. In order to ensure that there is enough power for one full treatment session, the battery icon must appear green.

Once sufficiently charged, remove the USB-C charging cable from the Monarch device.






Use of a partially-charged device may cause the Monarch to turn off before the treatment has completed.

Use only NeuroSigma approved charging cables and plugs. Use of unapproved accessories may cause loss of therapeutic effect, fire or explosion.

Clean the skin of the forehead with soap and water or an alcohol-based cleaning pad.


Completely dry the skin.



Remove electric patch from the release liner backing.

Using a mirror, center the patch on the forehead and apply just above the eyebrows.


Be sure the patch is securely and completely attached to the skin.






Connect electric patch wires to the lead wires.

Wires may be placed over and behind the ears.



Insert lead wire 3.5mm plug into the 3.5mm port at the bottom of the Monarch device.

Be sure the 3.5 mm plug is securely connected.





Treatment

Press and hold the ON/OFF button for 5 seconds to turn on. Display screen will illuminate and welcome you to the system.

Once all connections have been made, the home screen will appear and prompt you to press O to begin treatment. The initial level of stimulation will be 0.0mA. Increase the amplitude to a comfortable setting using the + button.

Press the UP key to increase treatment level to maximum tolerable level.
Stimulation should not be painful or uncomfortable.

Once the desired treatment level is set, you are ready to initiate the session by pressing the 0 button.

Move the lock slider into the “locked” position. In this state, all controls other than the lock slider will be disabled, ensuring no settings are changed inadvertently. To unlock, move the slider back to the unlocked position.

During sleep, the device can be placed under a pillow or on a bedside table.

Course of Treatment

Eight hours of continued nightly use is recommended. Consult your physician for any questions .

Clinical trials suggest that positive response may be seen within 4 weeks.

Once a clinical response is achieved continued therapy
may be necessary. Consult your physician.


In case of skin irritation, stop use and consult your physician.
Do not apply electric patches to broken, inflamed or infected skin.


Do not attempt to clean or reuse electric patches. Dispose after a single use.

⚠ A sudden, electric stimulus, bothersome to some users, may be experienced under the following conditions:

- When a lead wire is broken.
- When the contact at the following locations is poor:

Between the patch and the skin.

At the connection between the patch wire and the lead wire.

At the connection between the lead wire and the Monarch.

In the instance of a sudden electric stimulus, replace the lead wire and electric patch and restart treatment at a low current amplitude. If the sudden stimulus continues, stop using the Monarch and contact your physician immediately.

If the Monarch detects an interruption in any of the connections within the Monarch eTNS System, the message 'ensure all cables are attached correctly' will appear across the display screen and treatment will be paused. To resume function, make sure all the connections between the Monarch unit, the lead wires, and the patch are secure, then press 0 to resume treatment when prompted.

⚠ If you wake up and the Monarch is powered off, first make sure all the connections are intact, then, press and hold the ON/OFF button. If the Monarch does not turn on, it is likely out of battery. Charge the device via the USB cable provided for future use.

If there are recurrent device turn-offs, contact your physician.

After Treatment

When treatment is complete, unlock the Monarch by moving the lock slider to the unlocked position.

The screen should indicate a successful session. If there was an error or an interruption to the session, the display screen will tell you how long the treatment lasted before the interruption.

Turn the Monarch off by pressing and holding the ON/OFF button for 5 seconds. Display screen will turn off.

Unplug the lead wire from the Monarch.

Unplug the lead wire from the electric patch wire.

Gently remove the electric patch and dispose. Do not reuse.

Gently wash or clean skin of the forehead.

Cleaning Instructions

The Monarch device can be periodically cleaned with a cloth dampened with either water or a mild cleaning solution. Do not apply water or cleansers directly to any part of the Monarch eTNS System. Completely dry any component with a non-abrasive cloth before initiating treatment.

⚠ Do not clean the device while in use.

Do not make any alterations or repairs to the Monarch or to any of the components in the eTNS System, as this may result in personal injury, damage to equipment and voiding the manufacturer's warranty.

The Monarch may be damaged by dropping, shaking or rough handling, which may result in device malfunction, loss of treatment effect and personal injury.

Store the Monarch eTNS System in a dry environment avoiding excess humidity, temperature, sunlight and dust.

Do not store near chemicals or gases.

⚠ Cautions

The Monarch eTNS System generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause interference to other devices in the vicinity. However, there is no guarantee that interference will not occur in any particular installation. If this equipment does cause interference to other devices, try to correct the interference by one or more of the following measures:

– Reorient or relocate the receiving devices.

– Increase the separation between the equipment.

– Consult the manufacturer or field service technician for assistance.

– Electronic monitoring equipment (such as ECG monitors and ECG alarms) may not operate properly when eTNS stimulation is in use.

Note: There are no user-serviceable parts inside the unit. Do not attempt to modify or break open the device. If any service is required, please contact the manufacturer.

Environmental Specifications

Guidance and Manufacturer's Declaration - Electromagnetic Emissions

Medical electronic devices are designed to ensure electromagnetic compatibility (EMC).

Portable and mobile RF communications devices may affect medical electronic devices.

Do not place the Monarch next to or on top of another device when using it. If it has to be placed next to or on top of another device, check that this instrument and the device function properly before use. If accessories other than those supplied as spare parts by NeuroSigma are used, the emission of the Monarch may increase and immunity may be reduced.

The Monarch eTNS System is intended for use in the electromagnetic environment specified in the following tables. The customer or the user of the Monarch® should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic Environment— guidance
RF emissions CISPR 11	Group 1	The Monarch uses RF energy for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The Monarch is suitable for use in all establishments including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/flicker emissions IEC 61000-3-3	Complies	

Environmental Specifications (continued)

Guidance and Manufacturer's Declaration – Electromagnetic Immunity

Part 1

Immunity Test	IEC 60601-1-2 test level	Compliance level	Electromagnetic environment - guidance
Electromagnetic discharge (ESD) IEC 61000-4-2	± 6 kV contact ± 8 kV air		Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst IEC 61000-4-4	± 2 kV for power supply lines ± 1 kV per input/output lines		Not applicable. There is no connection to a.c. mains.
Surge IEC 61000-4-5	± 1 kV line(s) to line(s) ± 2 kV line(s) to earth		Not applicable. There is no connection to a.c. mains.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5% Ur (>95% dip in Ur) for 0.5 cycle 40% Ur (60% dip in Ur) for 5 cycles 70% Ur (30% dip in Ur) for 25 cycles <5% Ur (>95% dip in Ur) for 5 seconds		Not applicable. There is no connection to a.c. mains.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-6	3 A/m		Power frequency fields should be at levels characteristic of a typical location in a typical commercial, hospital or domestic envi- ronment.

NOTE: Ur is the a.c. mains voltage prior to application of the test level.

Environmental Specifications (continued)

Guidance and Manufacturer's Declaration – Electromagnetic Immunity

Part 2

Immunity Test	IEC 60601-1-2 test level	Compliance level	Electromagnetic environment - guidance
			Portable and mobile RF communications equipment should be used no closer to any part of the Monarch, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance:
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms	$d = 1.2 \sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m	$d = 1.2 \sqrt{P}$ 80 MHz to 800 MHz $d = 1.2 \sqrt{P}$ 800 MHz to 2.5 GHz Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,[a] should be less than the compliance level in each frequency range.[b] Interference may occur in the vicinity of equipment marked with the following symbol:

NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios,amature radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Monarch is used exceeds the applicable RF compliance level above, the Monarch should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the Monarch.

b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3V/m.

Recommended Separation Distances Between Portable and Mobile RF Communications Equipment and The Monarch eTNS System

The Monarch is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Monarch can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Monarch as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150 kHz to 80 MHz $d = 1.2 \sqrt{P}$	80 MHz to 800 MHz $d = 1.2 \sqrt{P}$	800 MHz to 2.5 GHz $d = 2.3 \sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	12	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance (d) in meters (m) can be estimated using the equation applicable
to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.

Electrical Specifications

Supply voltage:	Monarch unit 5 V DC
Input Ratings:	5 V DC 2.4A
Type of device:	Internally powered equipment/Type BF applied part
Dimensions:	40 mm (Width) x 112 mm (Length) x 13 mm (Depth)
Weight:	69 g
Display:	OLED
Service Life:	Monarch main unit: 18 months Components: Electric Patches: 1 year
Battery:	Li-ion Polymer, 1000mAh, integrated rechargeable battery
Accuracy of Measuring	All readings were taken between *1.0mA to 8.0mA *Error range: -5.5% at 1.0mA to -2.3% at 8.0mA
Functions:	Frequency: 120 Hz Pulse Width: 250µSec Load: 500Ω

Electrical Stimulation

Parameters	Specification
Intensity $I_{0\text{-}P}$(mA)	0 - 8 mA
Step (increment)	0.1 mA
Time	8 hours
Lock Mode	Locks all buttons except itself (Lock slider). Slide to unlock.

XII. Environmental Conditions

Operating environment: 10°C to 40°C, 30% to 75%, 700 to 1060 hPa



Storage environment: -10°C to 60°C, 30% to 95% (condensation allowed), 700 to 1060 hPa

Transport environment: -10°C to 60°C, 30% to 95% (condensation allowed), 700 to 1060 hPa


Clinical Data

Pilot Study

The pilot study was an 8-week, single arm, open label trial designed to assess the safety, tolerability, and initial efficacy of the Monarch eTNS. The subjects were aged 7 - 14 years old and were diagnosed with ADHD based on DSM-IV criteria.[1] Of the 24 subjects who initiated the study, 22 subjects completed the week 4 efficacy assessment and 21 completed the week 8 efficacy assessment, indicating a high level of tolerability for therapy with eTNS.

There were no serious adverse events associated with use of the device. Of the spontaneously reported adverse events, only headache (n = 2) and eye twitch (n = 1) were considered potentially device-related; all of these resolved spontaneously and did not inhibit trial participation. eTNS achieved a statistically significant improvement on the trial's primary endpoint, the ADHD-IV Rating Scale ("ADHD-RS"),[2] at weeks 4 and 8. Mean scores improved from 32.6 points at baseline to 18.2 points at Week 4 and 17.3 points at Week 8 ($p < 0.0001$).

Although not tested for statistical significance, assessment of subjects using the Clinical Global Impression – Improvement (CGI-I)[3] scale also demonstrated that a large majority of subjects improved in their overall clinical conduction. Specifically, at Week 4 and 8, 64% and 71% of subjects, respectively, were assessed as being "improved" or "very improved". In addition, according to daily treatment diaries from subjects who completed the trial, nightly treatment compliance was 100%.



Clinical Data (continued)

Pivotal Study

The pivotal study was a prospective, double-blind, randomized controlled trial of the Monarch eTNS as monotherapy for pediatric ADHD, with 4 weeks of treatment. The study enrolled 62 children (32 active; 30 sham) who had been diagnosed with moderated to severe ADHD.
Fifty-nine (59) subjects completed the four-week double-blind treatment period (31 active and 28 sham).

The average subject age was 10.4 years (range 8-13 years old). Sixty-five percent (65%) of subjects were male and thirty-five percent (35%) were female. The study population include White (65%), Black (6%), Asian (16%), and Mixed/Other (13%) races. The study population was 16% Hispanic. Sixty-three percent (63%) of subjects had the combined subtype of ADHD, thirty-four percent (34%) had the inattentive subtype, and three percent (3%) had the hyperactive/impulsive subtype.

The subjects received nightly treatments with either the active or sham device. The primary endpoint, as in the pilot study, was improvement on the clinician-administered ADHD-RS (weekly assessments). The CGI-I assessment was also administered.

[1] DSM-IV – Fourth edition of the Diagnostic and Statistical Manual of Mental Disorders. The DSM-IV is the primary assessment tool in the US for diagnosing ADHD.

[2] Attention Deficit Hyperactivity Disorder Rating Scale (ADHD-RS) – a parent rating scale that assesses the frequency of each ADHD symptoms identified in the DSM-IV criteria for the disease.


[3] Clinical Global Impression – Improvement (CGI-I): CGI-I is a physician rated assessment taken by interviewing the patients. The physician compares the patient's overall clinical condition to the week just prior to starting treatment and rates the comparison on a scale of 1 (very much improved) to 7 (very much worse). A score of 1-3 is associated with improvement, a score of 4 is no change, and a score of 5-7 is associated with deterioration.

Clinical Data (continued)

Pivotal Study (continued)

The trial met its primary efficacy endpoint. At the end of Week 4, the average ADHD-RS score in the active group decreased from 34.1 points at baseline to 23.4 points, versus a decrease from 33.7 to 27.5 points in the sham group. Using the GLMM methods for statistical analysis, it was found that the mean ADHD-RS scores showed differential group effects in favor of active eTNS as demonstrated by a significant group by time interaction (p = 0.005).

Mean ADHD-RS Total Score by Time Point by Group




The CGI-I scale results demonstrated a statistically significantly greater improvement in the active arm than the sham arm, when subjects were either classified as "improved" based on receiving a CGI-I score of 1 or 2 (i.e. "very improved" or "improved" on the assessment scale) or "not improved" based on a score ≥ 3. At Week 4, 52% of active subjects were classified as "improved" compared to only 14% of subjects in the sham arm (p = 0.003). Note, in the active group the percentage of subjects classified as "improved" rose each consecutive week, compared to the sham group which remained flat.

Clinical Data (continued)

eTNS was well-tolerated in all 62 subjects and there were no serious adverse events (SAEs) associated with use of the device. Commonly reported side effects among all pivotal study subjects included trouble sleeping, hyperactiveness, drowsiness, and stuffy nose, but there was no statistically significant difference between the active and sham groups for these events. Events that were more common in the active group than the sham group included: nightmares (6% versus 0%), headache (13% verus 0%), frequent urination (6% versus 0%), increased appetite (19% versus 7%), skin rash (6% versus 0%), and teeth clenching (13% versus 7%). The between group difference for these events was not statistically significant, except in the case of headache (p = 0.05). Weight gain was only slightly greater in the active group compared to the sham group (0.6 kg) and was not clinically meaningful given the age group. As noted above, no participant left the study due to side effects or adverse events in either the pilot or pivotal study. However, all reported side effects and adverse events were of low to moderate severity and most were transient.

Percent of Group Reporting Improvement on CGI-I Scale



Warranty

NeuroSigma, Inc. warrants the Monarch eTNS device against any defect in materials and workmanship for a period of two (2) years from the purchase date. Please retain all invoices and sales receipts, since these are proof of the purchase date.

The warranty does not cover accessories and spare parts, including electric patches, lead wires, and travel case. The warranty does not cover damage caused by rough handling, such as drops, shocks and exposure to water and other damaging environments.
The warranty does not apply to devices whose cases and/or screens are crushed, cracked, broken, or buckled. During the warranty period, at the sole discretion of NeuroSigma, the Monarch device will either be repaired free of charge by an approved service provider or replaced.

To obtain warranty service, go to our website, www.monarch-etns.com, and complete the Return Material Authorization (RMA) form and either mail or email it to our office, using the contact information provided on the website. You can also call our Customer Service Department at the following telephone number: +1.310.479.3100.

NOTE: All warranty repairs and service must be performed by NeuroSigma.

NOTE: A complete listing of all warranty features and exclusions is included as part of your purchase invoice. Please refer to it for further details or go to our website, www.monarch-etns.com, and download a full copy of the Monarch eTNS System warranty.

For more information about eTNS therapy and the side effects reported to date, please contact NeuroSigma.

Manufacturer
NeuroSigma
10960 Wilshire Boulevard, Suite 1910
Los Angeles, CA 90024 United States of America
Phone: +1.310.479.3100
www.neurosigma.com



Monarch and eTNS are trademarks of NeuroSigma, Inc. Patent pending.
© Copyright 2023 NeuroSigma, Inc., Los Angeles, CA, USA. All rights reserved.

760-0003-004 Rev B, 08/2026

NeuroSigma®, Inc.
10960 Wilshire Boulevard Suite 1910 Los Angeles, California 90024 USA